

## **CHLORIDE IN SOILS**

(An Arizona Method)

### **1. SCOPE**

- 1.1 This test method describes a procedure for determining chloride content in soil by a standard addition technique using a chloride electrode. The chloride content is defined in terms of the method and may be called water soluble chloride.
- 1.2 This test method involves hazardous material, operations, and equipment. This test method does not purport to address all of the safety concerns associated with its use. It is the responsibility of the user to consult and establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.
- 1.3 See Appendix A1 of the Materials Testing Manual for information regarding the procedure to be used for rounding numbers to the required degree of accuracy.
- 1.4 The extraction procedure, Subsections 4.1 through 4.4, is the same as is used in Arizona Test Method 733, "Sulfate In Soils".

### **2. APPARATUS**

- 2.1 Requirements for the frequency of equipment calibration and verification are found in Appendix A3 of the Materials Testing Manual. Apparatus shall consist of the following:
  - 2.1.1 Tall-form beakers - two, 200 mL each (Pyrex #1060 or equivalent), calibrated to indicate 100 mL volume.
  - 2.1.2 An analytical balance capable of measuring the maximum weight to be determined and conforming to the requirements of AASHTO M231, except the readability and sensitivity of any balance utilized shall be at least 0.001 gram.
  - 2.1.3 Magnetic stirrer and stirring bar.

- 2.1.4 Erlenmeyer flask(s) - 500 mL capacity with stopper(s).
- 2.1.5 Chloride electrode system - (Orion Ionplus #9617 combination chloride electrode or equivalent), suitable for use in a 200 mL tall-form beaker.
- 2.1.6 Specific ion meter - (Orion EA 940 or equivalent), suitable for use with the chloride electrode system.
- 2.1.7 Thermometer - accurate to at least 0.5 °C.
- 2.1.8 Pipettes - 1 mL, 5 mL, and 10 mL capacity, accurate to 1%.
- 2.1.9 Sample bottle - 200 mL capacity or larger, with cap.
- 2.1.10 Dropping bottle - 60 mL, with dispensing tip or dropper.
- 2.1.11 Centrifuge tube - 50 mL, with cap.

### 3. REAGENTS

- 3.1 Chloride Standard Solution, 1.000 mg/mL (1000 ppm).
  - 3.1.1 Fisher Scientific # LC13000-1 or equivalent. Alternatively, the Chloride Standard Solution may be made by transferring 1.648 gram dried primary standard sodium chloride into a 1000 mL volumetric flask and diluting to 1000 mL with demineralized water.
- 3.2 Nitric Acid, 20%.
  - 3.2.1 Measure 10 mL concentrated nitric acid into a suitable size beaker containing approximately 40 mL demineralized water and mix well. Transfer to a dropping bottle.

**CAUTION: Exercise extreme caution in preparing and using the Nitric Acid solution. It must be properly labeled and treated as a hazardous material.**

- 3.3 Demineralized Water.

#### 4. PROCEDURE

- 4.1 Weigh  $100.0 \pm 0.1$  grams of soil passing a No. 10 sieve into a 500 mL Erlenmeyer flask.
- 4.2 Weigh  $300.0 \pm 0.1$  grams demineralized water into the flask.
- 4.3 Stopper the flask, shake vigorously, and let the mixture stand undisturbed for one hour.
- 4.4 Carefully, with minimal disturbance of the sediment, decant 50 mL of the extract solution into a 50 mL centrifuge tube.
- 4.4.1 If the decanted extract solution is not cloudy, proceed to Subsection 4.5.
- Note:** The decanted extract solution does not have to be clear, as it is required to be in Arizona Test Method 733.
- 4.4.2 If the decanted extract solution is cloudy, place the tube into the centrifuge and centrifuge at a minimum of 5000 RPM for 10 minutes.
- 4.5 Pipette a 10.0 mL aliquot of the extract solution (or centrifugate) into a 200 mL beaker. Record this volume as "A". Dilute to the 100 mL mark with demineralized water and add three drops of Nitric Acid, 20%. This is the "Sample Reading Solution."
- 4.6 Place a stirring bar into the beaker, place the beaker onto the magnetic stirrer, insert the electrodes, and initiate stirring. Stirring shall be at a constant moderate rate, such that the vortex created by stirring does not expose the tips of the immersed electrodes. The rate of stirring shall be constant throughout the procedure.
- 4.7 After the reading has stabilized, record the initial reading to the nearest millivolt, as " $E_1$ ".
- 4.8 Add 1.00 mL of Chloride Standard Solution.
- 4.9 After the reading has stabilized, record the final reading to the nearest millivolt, as " $E_2$ ".
- 4.10 Calculate  $\Delta E$ . ( $\Delta E = E_1 - E_2$ ).

**Note:** " $\Delta E$ " is calculated and recorded initially as " $\Delta E_o$ " for the Sample Reading Solution. It is subsequently also calculated and recorded as " $\Delta E_b$ " for the Reagent Blank Reading Solution.

- 4.11 If " $\Delta E_o$ " (for the Sample Reading Solution) is less than 18, repeat Subsections 4.5 through 4.10 with a suitably smaller size aliquot, recording this aliquot volume as "A."
- 4.12 Prepare a Reagent Blank Reading Solution by placing 100 mL of demineralized water into a second 200 mL beaker and adding 3 drops of Nitric Acid 20%. Repeat Subsections 4.6 through 4.10, calculating and recording " $\Delta E_b$ " (for the Reagent Blank Reading Solution).

## 5. CALCULATIONS AND REPORT

- 5.1 Calculate chloride concentration, " $C_o$ " in the Sample Reading Solution and " $C_b$ " in the Reagent Blank Reading Solution, and record each to the nearest 0.001 mg/mL, as follows:

$$C_o = \frac{1}{\left[ 101 \times 10^{[(\Delta E_o)/S]} \right] - 100}$$

$$C_b = \frac{1}{\left[ 101 \times 10^{[(\Delta E_b)/S]} \right] - 100}$$

Where:

S = Electrode slope, as determined in accordance with the manufacturer's recommendations. (The slope should equal approximately 59 millivolts for a properly functioning electrode.)

- 5.2 Calculate chloride concentration in the soil in parts per million, "C", and report to the nearest 10 ppm as follows:

$$C = \frac{300,000 \times (C_o - C_b)}{A}$$